


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-5B-C-CA-P*

S.N.: *5731*

Date: *DEC 10 1990*



*R<sub>L</sub> = 0.7  $\Omega$*   
*10. VOLTS/DIV*  
*100 NS/DIV*  
*PRF =*

- a) Output signal amplitude:  
*0 TO +48 Amps*
- b) Pulse width: *(to 0.7  $\mu$ s)*  
*25 TO 500 NS*
- c) Rise time:  
*< 10 NS*
- d) Fall time:  
*< 10 NS*
- e) PRF:  
*0 TO 1KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240V*  
*50-60 Hz*  
*[Signature]*