


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V-1A2-P-CGA*

S.N.: *5732*

Date: *JAN 7 1991*



60 db ATTEN

100 VOLTS DIV

PRF = 20 KHz

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
0 TO 20 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 VOLT

250 mA (MAX)