


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-5A-C-P-077A*

S.N.: *5733*

Date: *DEC 12 1990*



*10 VOLTS/DIV*  
*R<sub>i</sub> = 0.7  $\Omega$*   
*100 NS/DIV*  
*PRF = 5.0 KHz*

- a) Output signal amplitude:  
*0 TO +40 Amps*
  - b) Pulse width: *TO 0.7  $\mu$*   
*20 TO 250 NS*
  - c) Rise time:  
 *$\leq 5$  NS*
  - d) Fall time:  
 *$\leq 5$  NS*
  - e) PRF: *0 TO 5 KHz*
  - f) Jitter, stability:  
*OK*
  - g) Prime power:  
*120/240 V*  
*50-60 Hz*
- [Signature]*