


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVS-3B-C-D*
S.N.: *5734*
Date: *DEC 12 1990*



- a) Output signal amplitude: *+10 TO +30 Amp*
- b) Pulse width: *≤ 1.5 ns*
- c) Rise time: *(HALF period) NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 20 kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120 / 240V, 50-60 Hz.*

R_c = 5.0 Ω
20 Amps/DIU
PRF = 20 kHz

[Signature]