


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVI-P
S.N.: 5737
Date: JAN 7 1991



- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width:
2 TO 100 NS
- c) Rise time:
≤ 350 PS
- d) Fall time:
≤ 2 NS
- e) PRF:
MIN PW: 0 TO 1 MHz
MAX PW: 0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 V,
150 mA (MAX)

50 dB ATTEN
32 VOLTS/DIV
100 KHz

[Handwritten signature]