


PULSE GENERATOR
PERFORMANCE CHECK

Model: *NPP-1-C-M-T*
S.N.: *5741*
Date: *DEC 18 1990*



A RANGE OUT
40 dB ATTEN
10 VOLT/DIV
PRF = 1.0 MHz



B RANGE OUT
40 dB ATTEN
10 VOLT/DIV
PRF = 1.0 MHz

- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
0.2 TO 100 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
170/240 V
50-60 Hz

[Signature]