


PULSE GENERATOR
PERFORMANCE CHECK


Model: *PVL-17T5-M*

S.N.: *5742*


Date: *DEC 7 1990*



① MAIN FRAMES OUT TO 50Ω
60 dB ATTEN
100 VORTS/DIV



② 17T5-M1 OUT TO
1.5 K PROBE (TO 40 dB)



③ MONITOR OUT FOR
② (TO 40 dB)

- a) Output signal amplitude:
-200 V TO AC > 1 K.
- b) Pulse width:
5 TO 200 NS
- c) Rise time:
≤ 3 NS
- d) Fall time:
≤ 3 NS
- e) PRF:
0 TO 1 K/RZ
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240 V
50-60 ITR