



PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVP-AV-2-PS-P*  
 S.N.: *5751*  
 Date: *DEC 17 1990*



- a) Output signal amplitude:  
*0 TO +10 VOLTS*
- b) Pulse width:  
*5 TO 50 NS*
- c) Rise time:  
*≤ 100 PS*
- d) Fall time:  
*≤ 135 PS*
- e) PRF:  
*0 TO 50 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*

*Ⓟ 40 dB ATTEN*  
*10 VOLTS/DIV*  
*PRF = 20 KHz*



*Ⓟ 100 dB ATTEN*  
*1 NS/DIV*  
*(RISE TIME)*