


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PAVA-1122-2-N*  
S.N.: *5753*  
Date: *DEC 17 1990*



*50 66 ATTEN*  
*32 VOLTS/DIV*  
*PRF = 500 kHz*

- a) Output signal amplitude:  
*≥ -60 VOLTS*
- b) Pulse width:  
*≤ 2 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*135 ± 5 VOLTS*  
*5 mA max*

*[Signature]*