



PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PVL-3A-PS-N*  
S.N.: *5761*  
Date: *MAR 14 1991*



Point *70 dB ATTEN*  
*320 VOLTS/DIV*



Norm *(AS ABOVE)*

- a) Output signal amplitude:  
*0 TO ± 300 VOLTS*
- b) Pulse width:  
*5 TO 100 NS*
- c) Rise time:  
*≤ 600 PS*
- d) Fall time:  
*≤ 3 NS*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*  
*50-60 Hz*

*[Signature]*