


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVP-AN-15-C-N*


S.N.: *5763*

Date: *JUNE 12 1991*



- a) Output signal amplitude:  
*0 TO -5 VOLTS*
- b) Pulse width:  
*0.2 TO 4.0 NS*
- c) Rise time:  
*≤ 60 PS*
- d) Fall time:  
*≤ 135 PS*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*

*(A) 40dB ATTEN*  
*5 VOLTS/DIV*  
*PRK = 100 KHz*



*(B) AS (A) BUT 100 PS/DIV*  
*(RISE TIME)*