


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PMU-EA-C-7*

S.N.: *5764*

Date: *FEB 26 1991*



60 db ATTEN

100 VOLTS/DIV

PRF = 100KHZ

a) Output signal amplitude:
0 TO 700 VOLTS

b) Pulse width:
1 TO 5 NS

c) Rise time:
≤ 0.4 NS

d) Fall time:
≤ 0.4 NS

e) PRF:
0 TO 100 KHZ

f) Jitter, stability:
OK

g) Prime power:
120 / 240 V
50-60 HZ

[Signature]