


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVI-U-PS-P*

S.N.: *5771*

Date: *FEB 8 1991*



Ⓐ *50 dB ATTEN*  
*≈ 32 VOLT/DIV*  
*PRF = 10 KHz*



Ⓑ *175 @ OUT*  
*1 NS/DIV*  
*(RISE TIME)*

- a) Output signal amplitude:  
*0 TO + 50 VOLTS*
- b) Pulse width:  
*3.7 NS*
- c) Rise time:  
*≤ 350 PS*
- d) Fall time:  
*≤ 2 NS*
- e) PRF:  
*0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240 V*  
*50-60 Hz*  
*[Signature]*