



PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-5B-T*  
 S.N.: *5773*  
 Date: *JAN 22 1991*



- a) Output signal amplitude:  
*0 TO 48 AMP TO*
- b) Pulse width: *0.75  $\mu$ s*  
*40 TO 500 NS*
- c) Rise time:  
 *$\leq 10$  NS*
- d) Fall time:  
 *$\leq 10$  NS*
- e) PRF: *0 TO 1KHz*
- f) Jitter, stability:  
*NA*
- g) Prime power:  
*NA*

(A)  
OUTPUT: *10 VOLT/DIV*  
*100 NS/DIV*  
 *$R_L = 0.75 \mu$*   
 *$I \approx 48$  AMP*



*A, B. BOTH*  
*DRIVEN BY*  
*5773-10*  
*AVO-5B-C*  
*MAINTENANCE*  
*[Signature]*

(B)  
OUTPUT: *10 VOLT/DIV*  
*100 NS/DIV*  
 *$R_L = 0.75 \mu$*   
 *$I \approx 20$  AMP*