


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMMP-3-C-P*


S.N.: *5774*

Date: *JAN 30 1991*



- a) Output signal amplitude:
0 TO +20 VOLTS
- b) Pulse width:
8 TO 100 NS
- c) Rise time:
≤ 200 PS
- d) Fall time:
≤ 300 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz.

Ⓐ *40 dB ATTEN*
10 VOLTS/DIV
PRF = 1.0 MHz



Ⓑ *AS Ⓐ BUT*
1.0 NS/DIV.