


PULSE GENERATOR

PERFORMANCE CHECK


Model: *ANPP-1-C-P-M*

S.N.: *5776*

Date: *Feb 25 1991*



*A out 40db ATTEN
-10 VOLTS/DIV
PRF = 1.0 MHz*



*B out 40 db ATTEN
-10 VOLTS/DIV
PRF = 1.0 MHz*

- a) Output signal amplitude:
A) 0 TO 710 VOLTS
- b) Pulse width:
B) 0 TO 470 VOLTS
A) 0.2 TO 5.0 NS
- c) Rise time:
B) 5.0 TO 100 NS
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz