


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMR-B4-C-P*
S.N.: *5777*
Date: *FEB 25 1991*



- a) Output signal amplitude:
0 TO 7700 VOLTS
- b) Pulse width:
1 TO 5 NS
- c) Rise time:
≤ 0.4 NS
- d) Fall time:
≤ 0.4 NS
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240V
50-60 Hz.

60 db ATTEN
~ 700 VOLTS DIV
PRF = 100 KHz

[Signature]