


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMP-M-1-C-P-SLAC1*

S.N.: *5781*

Date: *FEB 15 1991*



*40 dB ATTEN*

*10 DIVS/DIV*

- a) Output signal amplitude:  
*0 TO +1000mV*
- b) Pulse width: (A+B)  
*0.5 TO 7.0 nSEC*
- c) Rise time: (A+B)  
*≤ 500 PS*
- d) Fall time:  
*≤ 500 PS*
- e) PRF:  
*0 TO 100 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 V*  
*50 - 60 Hz*
- h) T: *0 TO 200 nS.*