


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-A-1-PW-C-CA-P*

S.N.: *5786*

Date: *FEB 13 1991*



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width:
50 NS TO 100 NS
- c) Rise time: *(0.5% MAX DUTY CYCLE)*
≤ 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF: *0 TO 100 KHz*
0.5% MAX DUTY CYCLE
- f) Jitter, stability:
OK
- g) Prime power:

R_L = 50 Ω
50 VOLTS / DIV
100 NS / DIV
PRF = 10 KHz

120 / 240 V
50 - 60 Hz