


PULSE GENERATOR
PERFORMANCE CHECK


Model: APR-4B-PW-C-D-TECH1

S.N.: 5787

Date: APRIL 2 1991



Ⓐ MIN PW
 $R_L = 10K$
(NO OFFSET)



Ⓑ MAX PW
 $R_L = 10K$
PRF = 1KHz

- a) Output signal amplitude:
0 TO +400 VOLTS
- b) Pulse width: TO $R_L > 10K$
50 NS TO 10 μS
- c) Rise time:
≤ 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF: 0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Handwritten signature]