


PULSE GENERATOR
PERFORMANCE CHECK


Model: *PP-101-422-C-N-PN-M-05-CIA*

S.N.: *5789*

Date: *FEB 5 1991*



*(A) OUT 40 db ATTEN
10 VOLT/DIV
PRF = 100 KHz*



*B OUT 40 db ATTEN
10 VOLT/DIV
PRF = 100 KHz*

- a) Output signal amplitude:
0 TO ± 20 VOLTS
- b) Pulse width:
0.3 TO 2.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 200 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*120/240 V
50-60 Hz*
- h) DEBURS JTB:
0 TO ± 6 NS

[Signature]