


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMR-B3-W-C-P-GA*

S.N.: *5790*

Date: *FEB 7 1991*



- a) Output signal amplitude: *0 TO +100 VOLTS*
- b) Pulse width: *TO 50~*
- c) Rise time: *0.1 TO 5.0 NS*
- d) Fall time: *≤ 5 NS*
- e) PRF: *0 TO 5 KHz.*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240V*

R_L = 50Ω
20 VOLTS/DIV
0.2 NS/DIV.
PRF = 5 KHz

120/240V
50-60 Hz.
[Signature]