


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ACL-2A-C-P*

S.N.: *5794*

Date: *FEB 12 91*



*60 db ATTEN*

*100 VOLTS/DIV*

*PRF = 5.0 KHZ*

- a) Output signal amplitude:  
*0 TO +170 VOLTS*
  - b) Pulse width:  
*5 TO 100 NS*
  - c) Rise time:  
*< 2 NS*
  - d) Fall time:  
*< 3 NS*
  - e) PRF:  
*0 TO 5 KHZ*
  - f) Jitter, stability:  
*OK*
  - g) Prime power:  
*120/240 V*  
*50-60 KHZ*
- [Signature]*