



PULSE GENERATOR
PERFORMANCE CHECK

Model: *APK-AN-C-PN*
S.N.: *5796*
Date: *MAR 8 1991*



(A) *40dB ATTEN*
10 VOLT/DIV



(B) *AS (A) BUT 1.0 NS/DIV*
(RISE TIME)

- a) Output signal amplitude:
0 TO ±15 V_{LOAD}
- b) Pulse width:
1-50 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 50 K/HZ
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 HZ