


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVG-1-P*
S.N.: *5797*
Date: *MAR 4 1991*



60 dB ATTEN
-100 VOLTS DIV
PRF = 50 KHz

- a) Output signal amplitude:
0 TO +250 VOLTS
- b) Pulse width:
≤ 4 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS (± 1 VOLT)
200 mA (MAX)

[Signature]