


PULSE GENERATOR  
PERFORMANCE CHECK


Model: *PAK-32-PN-C-MOTAR*

S.N.: *5799*

Date: *MAR 26 1991*



*Point 50 dB ATTEN  
∴ 32 VOLTS/DIV  
PRF = 100 KHz*



*Now 50 dB ATTEN  
∴ 32 VOLTS/DIV  
PRF = 10 KHz  
PW = 1.0 μs  
(RISE TIME)*

- a) Output signal amplitude:  
*0 TO ± 50 VOLTS*
- b) Pulse width:  
*200 NS TO 1.0 μs  
(2% MAX DUTY CYCLE)*
- c) Rise time:  
*≤ 2 NS*
- d) Fall time:  
*≤ 2 NS*
- e) PRF:  
*0 TO 100 KHz  
(2% MAX DUTY CYCLE)*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V  
50-60 Hz.*

*[Signature]*