


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MH-5-1-C-05-P*

S.N.: *5803*

Date: *MAR 18 1991*



40 db ATTEN
10 VOLTS DIV
PRF = 100 KHz

- a) Output signal amplitude:
0 TO 110 VOLTS
- b) Pulse width:
≤ 130 ps
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240V
50-60 Hz