


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-AV-1-C-P-05-W*

S.N.: *5810*

Date: *APRIL 16 1991*



60 db ATTEN

~ 100 VOLTS/DIV

PRF = 50 KHz

- a) Output signal amplitude: *0 TO 100 VOLTS*
- b) Pulse width: *5 TO 400 NS**
- c) Rise time: *≤ 1 μS*
- d) Fall time: *≤ 3 μS*
- e) PRF: *0 TO 5 KHz*
- f) Jitter, stability: *O.K.*
- g) Prime power: *120/240 V*
50, 60 Hz

** FOR 100 TO 400 NS
RANGE NEEDS
ADDITIONAL
EXTERNAL DEBIT
LINE*

[Signature]