

PULSE GENERATOR

PERFORMANCE CHECK


Model:

AVO-5A-P-M1-TTL


S.N.:

5812

Date:



Ⓐ 60 db ATTEN
 100 VOLTS/DIV
 100 AMP/5/DIV
 PRF = 7 KHZ
 (MIN PW)




Ⓑ AS Ⓐ BUT
 WITH PW

- a) Output signal amplitude:
0 TO +40 AMP(?)
- b) Pulse width: LT30-24 IN
SCRIBED WITH 1
- c) Rise time: 5 TO 20 NS (HARF POW)
NA
- d) Fall time: NA
- e) PRF: 0 TO 8 KHZ
- f) Jitter, stability: OK
- g) Prime power:

+15 VOLTS,
 400 mA (MAX)

- h) PW CONTROL
 0 TO +10 VOLTS
 DC (R_{in} > 10K)



[Signature]