


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-1-C-P*
S.N.: *5813*
Date: *FEB 28 1991*



ca 60 Amps / DIN
PRF = 1.0 kHz

- a) Output signal amplitude:
+25 to +100 Amps
- b) Pulse width:
≤ 25 ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 1 kHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

[Signature]