


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVD-3A-C7*  
S.N.: *5814*  
Date: *FEB 28 1991*



- a) Output signal amplitude:  
*+25 TO +75 AMPS*
- b) Pulse width:  
*2.5 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*$R_c = 4\Omega$*   
*60 db ATTEN*  
*100 VOLTS/DIV*  
*25 AMP/DIV*

*120/240 V*  
*50-60 Hz*