


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-A-1-PW-PS-05-P*

S.N.: *5615*

Date: *March 5 1991*



$V_c = 50 \sim$
 $PRF \approx 5 \text{ kHz}$

- a) Output signal amplitude:
0 to +200 volts
- b) Pulse width:
50 ns to 1.0 us
- c) Rise time: *0.5% max*
DUTY cycle:
≤ 10 ns
- d) Fall time:
≤ 10 ns
- e) PRF:
0 to 100 kHz
- f) Jitter, stability: *0.5% max*
DUTY cycle:
OK
- g) Prime power:
120/240V
50-60 Hz