


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-A-1-pw-ps-05-7*

S.N.: *5816*

Date: *APR 5 1991*



- a) Output signal amplitude:  
*0 TO +200 VOLTS*
- b) Pulse width:  
*50 NS TO 1.0  $\mu$ S*
- c) Rise time: *0.5% MAX DUTY CYCLE*
- d) Fall time:  *$\leq 10$  NS*
- e) PRF: *0 TO 100 KHZ*
- f) Jitter, stability: *0.5% MAX DUTY CYCLE*
- g) Prime power: *OK*

*$R_L = 50 \Omega$*   
*PRF = 5.0 KHZ*

*120/240 V*  
*50-60 HZ*

*M*