


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVMM-3-C-TR-P-PW*

S.N.: *5823*

Date: *MAR 22 1991*



a) Output signal amplitude:
0 TO ± 10 VOLTS

b) Pulse width:
1 TO 5 NS

c) Rise time:
< 0.3 TO 2.0 NS


d) Fall time:
< 0.6 TO 2.0 NS

e) PRF:
0 TO 25 MHz

f) Jitter, stability:
OK

g) Prime power:
120/200V
50-60 Hz

① *40 db ATTEN*
10 VOLTS/10V
PRF ≈ 5.0 MHz
MIN TR



② *40 db ATTEN*
10 VOLTS/10V
PRF ≈ 12 MHz
MIN TR