


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-3-PW-C-P*

S.N.: *5824*

Date: *MAR 11 1991*



- a) Output signal amplitude: *0 TO + 200 Volts*
- b) Pulse width: *TO 50 μ s*
0.1 TO 100 μ s
- c) Rise time: *≤ 10 ns*
- d) Fall time: *≤ 10 ns*
- e) PRF: *0 TO 1 kHz*
(MAX PW, 0 TO 5 kHz)
- f) Jitter, stability: *OK*
- g) Prime power:

R_L = 50 Ω
30 Volts / DIV
2.0 μ s / DIV

120 / 240 V
50 - 60 Hz