


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-S-1-C-05-N

S.N.: 5828

Date: MAR 18 1991



40 dB ATTEN

10 V/DIV

PRF = 100 KHz

- a) Output signal amplitude:
0 TO -10 VOLTS
- b) Pulse width:
≤ 130 ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V
50-60 Hz

[Signature]