


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-3A-P-LLLA*


S.N.: *5833*

Date: *APRIL 9 1991*



- a) Output signal amplitude:
7.1 ± 300 VOLTS
- b) Pulse width:
10 NS
- c) Rise time:
< 600 PSEC
- d) Fall time: *(≈ 400 PSEC)*
≈ 4 NS
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
15 ± 1.0 VOLTS
250 mA
(MAX)

① *70 db ATTEN*
320 VOLTS/DIV
PRF = 10 KHZ



② *100 db ATTEN*
1 NS/DIV (RISE TIME)

[Signature]