


PULSE GENERATOR
PERFORMANCE CHECK


Model: *19VL-3A-P-LLCA*

S.N.: *5835*

Date: *APRIL 9 1991*



Ⓐ 70 Ω ATTEN
∴ 320 VOLTS/DIV
PRF = 10 KHz



Ⓑ AS Ⓐ BUT
1 NS/DIV (RISE TIME)

a) Output signal amplitude:

7, +300 VOLTS

b) Pulse width:

10 NS

c) Rise time:

< 600 PS

d) Fall time:

(≈ 400 PS)

≈ 4 NS

e) PRF:

0 TO 20 KHz

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 VOLTS

250 mA (MAX)