


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AP4-510-N*

S.N.: *5841*

Date: *APRIL 5 1991*



*40dB ATTEN*  
*10 VOLTS/DIV*  
*PRF = 700 kHz*

- a) Output signal amplitude:  
*0 TO 10 VOLTS*
- b) Pulse width:  
*≤ 130 ps*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240V*

*120/240V*  
*[Signature]*


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVH-S-1-C-N*

S.N.: *5841 (mod)*

Date: *JUNE 25 91*



*ⓐ 40 db ATTEN  
10 0000 / DIV  
PRF = 1.0 MHz*



*ⓑ AS ⓐ BUT  
100 PS / DIV*

- a) Output signal amplitude: *0 TO -10 VOLTS.*
- b) Pulse width: *200PS (TO 1.0 NS)*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V  
50-60 HZ*

*✓*