


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-E3-C-PN-05-W-CA-GW*


S.N.: *5843*

Date: *MAY 17 1991*




- a) Output signal amplitude:
0 TO ± 100 VOLTS
- b) Pulse width: *(TO 50ns)*
0.1 TO 5.0 us
- c) Rise time:
≤ 1 ns
- d) Fall time:
≤ 1 ns
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50 - 60 Hz


① *Post Ri = 50Ω PRF = 5 KHz*



② *Post Ri = 50Ω PRF = 5 KHz*



③ *Post 60db ATTEN*
∴ 100 VOLTS/DIV
(RISE TIME)



④ *Post 60 db ATTEN*
∴ 100 VOLTS/DIV
(RISE TIME)

[Signature]