


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-3A-C-10*

S.N.: *5846*

Date: *APRIL 19 1991*



≈ 25 AMP/DIV

PRF = 5 KHz

- a) Output signal amplitude:
± 25 TO ± 75 AMP
- b) Pulse width:
2.5 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

120/240 V

50-60 Hz

[Signature]