


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *A20-3B-C-P*  
S.N.: *5847*  
Date: *APRIL 19 1991*



*≈ 25 AMP/DIV*  
*PRF = 10 KHz*

- a) Output signal amplitude:  
*+ 10 TO 30 AMPS*
- b) Pulse width:  
*1.5 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*2NA*
- e) PRF:  
*0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V*  
*50-60 Hz*