


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *ANR-1-PW-75-T*

S.N.: *5850*

Date: *APRIL 23 1991*



- a) Output signal amplitude: *0 TO +200 VOLTS*
- b) Pulse width: *0.1 TO 5.0  $\mu$ S*
- c) Rise time:  *$\leq 10$  NS*
- d) Fall time:  *$\leq 10$  NS*
- e) PRF: *0 TO 1 K/Hz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V*  
*50-60 Hz*

*50 VOLTS/DIV*  
*0.5  $\mu$ S/DIV*  
 *$R_L = 50 \Omega$*   
*PRF = 1.0 KHz*

*[Signature]*