


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV-107B-C-P*

S.N.: *5851*

Date: *APRIL 29 1991*



$R_L = 22 \Omega$

$V = 20 \text{ VOLTS/DIV}$

$I = 1 \text{ AMP/DIV}$

100 NS/DIV

$\text{PRF} = 20.0 \text{ KHz}$

- a) Output signal amplitude:
0 TO +2 AMPS TO
- b) Pulse width: *0 TO +60 VOLTS*
20 TO 200 NS
- c) Rise time:
 $\leq 10 \text{ NS}$
- d) Fall time:
 $\leq 10 \text{ NS}$
- e) PRF: *0 TO 20 KHz*
- f) Jitter, stability:
OK
- g) Prime power:

$120/240 \text{ V}$

$50-60 \text{ Hz}$