


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MVN-1-C-N*

S.N.: *5854*

Date: *JUNE 24 1991*



*40 db ATTEN*

*5 VOLTS/DIV*

*PRFN 160 MHz*

- a) Output signal amplitude:  
*0 TO 5 VOLTS*
- b) Pulse width:  
*0.2 TO 1.0 NS*
- c) Rise time:  
*≤ 100 PS*
- d) Fall time:  
*≤ 100 PS*
- e) PRF:  
*50 TO 250 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240 V*  
*50-60 Hz*