


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-7C-N-M-AT-NR41*


S.N.: *5856*

Date: *MAY 10 1991*



- a) Output signal amplitude:
0 TO -30 Amps
- b) Pulse width:
TO 0 TO 30 VOLTS
0.2 TO 50 μ S
- c) Rise time:
 ≤ 50 NS
- d) Fall time:
 ≤ 50 NS
- e) PRF:
0 TO 5 KHz
(FOR PW ≤ 0.5 μ S)
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

Ⓐ *10 μ S/DIV*
10 AMP/DIV
PRF = 50 KHz



Ⓑ *10 μ S/DIV*
10 AMP/DIV
PRF = 2 KHz

[Signature]