


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVG-107*  
S.N.: *5862*  
Date: *APRIL 16 1991*



*60 db return*  
*.100 VOLTS/DIV*  
*PRF = 20 KHz*

- a) Output signal amplitude:  
*0 TO +250 VOLTS*
- b) Pulse width:  
*≤ 4 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 50 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*15 ± 1.0 VOLTS*  
*300 mA (MAX)*

*[Signature]*