


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-2-C-PN-05

S.N.: 5063

Date: MAY 14 1991



70 db ATTEN

$\approx 320$  VOLTS/DIV

PRF  $\approx 2.0$  KHz

- a) Output signal amplitude:  
0 TO  $\pm 350$  VOLTS
- b) Pulse width:  
5 TO 100 NS
- c) Rise time:  
 $\leq 2$  NS
- d) Fall time:  
 $\leq 3$  NS
- e) PRF:  
0 TO 5 KHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240 V  
50-60 Hz  
A