


PULSE GENERATOR

PERFORMANCE CHECK


Model: *MPP-2-PS-M-P*

S.N.: *5866*

Date: *MAY 15 1991*



A OUT 40 dB ATTEN
- 10 VOLTS/DIV
PRF = 50 kHz



B OUT 40 dB ATTEN
10 VOLTS/DIV
PRF = 50 kHz

- a) Output signal amplitude:
0 TO + 20 VOLTS
- b) Pulse width:
0.4 TO 100 NS
- c) Rise time:
≤ 200 PS
- d) Fall time:
≤ 300 PS
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
120 / 240 V
50 - 60 MHz

[Signature]