


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-1-PW-C-P

S.N.: 5867

Date: APRIL 26 1991



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width:
0.1 TO 5.0 μ S
- c) Rise time:
 ≤ 10 NS
- d) Fall time:
 ≤ 10 NS
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

50 VOLTS/DIV
0.5 μ S/DIV
 $R_L = 50 \Omega$
PRF = 1 KHz

120 / 240 V
50 - 60 Hz
[Signature]